

**Mapping pigmentation in human skin
from multi-channel visible spectrum image
by inverse optical scattering technique**

Norimichi Tsumura, Miki Kawabuchi,
Hideaki Haneishi, Yoichi Miyake

Department of Information and Image Sciences,
Chiba University, Japan

Abstract

Mapping pigmentation in human skin is expected to give useful information in reproducing skin color and enhancing the ability to diagnose various skin disease. In this research, maps of melanin, oxy-hemoglobin and deoxy-hemoglobin in skin are estimated from multi-channel visible spectrum image by using an inverse optical scattering technique. In the inverse optical scattering technique, first of all, a forward model of optical scattering is build to simulate the spectral reflectance of skin. Changing the variable parameters in the forward model, the simulation is repeated until the simulated spectral reflectance matches with the spectral reflectance at each pixel of the multi-spectral image. The principle of the proposed estimation technique was confirmed by imaging the human forearm under the venous occlusion, the venous and arterial occlusion, and by imaging a slapped region of the human forearm.

Keywords:

skin color, melanin, oxy-hemoglobin, deoxy-hemoglobin,
multi-spectral image, multi-channel visible spectrum image,
inverse optical scattering, Monte Carlo simulation, oxygen saturation,
venous and arterial occlusions

Introduction

Extracting qualitative information and spatial distribution of components such as pigmentation through analyzing an observed image is required in many fields of image analysis such as remote sensing, medical diagnostics, and robot vision. Spectral characteristics of the components are useful information to identify the components, because different materials have different spectroscopic responses to electromagnetic waves of a certain energy band. A technique that uses a spectral image as an observed image was proposed¹ and has been expected to lead the field of image analysis.

Skin color reproduction may be considered the most important problem in color reproduction of color film and color television systems. With the recent progress of various imaging systems²⁻⁴ such as multimedia, computer graphic and telemedicine systems, skin color becomes increasingly important for communication, image reproduction on hardcopy and softcopy, medical diagnosis, cosmetic development and so on. Human skin is a turbid medium with multi-layered structure^{5,6}, and various pigments such as melanin and hemoglobin are contained in the medium. Slight changes in structure and pigmentation produce rich skin color variation. Therefore, it is necessary to analyze skin color on the basis of the structure and pigmentation in reproducing skin color and enhancing the ability to diagnose various skin disease.

We have already proposed a technique to extract the map of skin pigmentation from skin color images⁷ or skin spectral absorbance image⁸ by independent component analysis. The facial color images were separated into the images that correspond to distribution of melanin and hemoglobin. The separated components were synthesized to simulate the various facial color images by changing the quantities of the two

separated pigments. In the analysis of skin color, it was assumed that the linearity among the quantities of pigment and the observed color signals holds in the optical density domain. However, the optical scattering in skin which causes the broad distribution of optical path length histogram in the skin may violate the linearity. It was also assumed that spectral absorbance of the pigment will not change spatially. However, hemoglobin has two type of state: oxy-hemoglobin(HbO_2), and deoxy-hemoglobin(Hb), and ratio between HbO_2 and Hb will change spatially in a large area of skin image or in an area of skin diseases. The oxygen saturation is defined by the ratio between oxy-hemoglobin and total hemoglobin. Mapping oxygen saturation of blood in skin will give useful information for skin diagnosis⁹.

In this paper, the pigmentations of melanin, oxy-hemoglobin and deoxy-hemoglobin in skin are estimated from multi-channel visible spectrum image by using the inverse optical scattering technique. The reflected visible light will provide information only for epidermis and dermis layers in the skin, because most of visible light does not penetrate into subcutis layer.

In the inverse optical scattering technique, first of all, a forward model of optical scattering is build to simulate the spectral reflectance of the skin. Changing the variable parameters in the forward model, the simulation is repeated until the simulated spectral reflectance matches with the spectral reflectance at each pixel of the multi-spectral image. The conventional non-linear optimization technique is used to change the variable parameters at each iteration.

In the next section, oxygen saturation is briefly reviewed which is the important diagnostic factor during and after operation, and relation between oxygen saturation and skin color is described. Next, the forward model of optical scattering in human skin is described. The inverse

optical scattering technique is explained based on the forward model. Finally, the experimental results of imaging the human forearm under venous occlusion, venous and arterial occlusion, and imaging a slapped region of the human forearm are shown.

Oxygen saturation

The oxygen saturation S_{O_2} is generally defined by the ratio between densities of oxy-hemoglobin and total hemoglobin as follows.

$$S_{O_2} = [HbO_2] / ([Hb] + [HbO_2]) \quad (1)$$

where $[Hb]$, $[HbO_2]$ indicate the densities of deoxy-hemoglobin and oxy-hemoglobin respectively. The oxygen saturation of the arterial blood in patient is usually monitored during and after the operation by pulse oximeter, because hypoxia that means lack of oxygen in the body is very dangerous during and after anesthesia in operation. The oxygen saturation 0.97 ~ 0.98 is normal in the arterial blood, about 0.70 is normal in venous blood.

The pulse oximeter analyzes the optical density of the transmitted near infrared light at the fingertip¹⁰. It is necessary to be concerned only with the attenuation of light by arterial blood in the total attenuation caused by arterial blood, venous blood, skin, muscle, bone, connective tissue and so on. It can be assumed that the arterial blood only changes its volume as the blood flows into and out of the vascular bed with pulsation. Therefore, it is possible to calculate the oxygen saturation of the arterial blood by subtracting the d. c. component of the attenuation from the total attenuation by the fingertip.

The near infrared light used in pulse oximeter can not be used to detect the pigmentation in skin layers, because most of the near infrared light transmit the skin. On the other hand, most of visible light is diffusely reflected by skin in epidermis and dermis layers because skin layers have strong scattering property for visible light. Therefore, the skin color is the light that is diffusely reflected by epidermis and dermis layers with absorptions by pigments such as oxy-hemoglobin, deoxy-hemoglobin, and melanin. It will be effective to reproduce skin color based on the pigmentations considering the oxygen saturation which reflect the human physical and mental conditions. The oxygen saturation in skin layers is also expected to give useful information for diagnosing skin diseases such as psoriasis⁹. It is noted that pulsation can not be used in skin layers, and strong scattering also make this measurement difficult without modeling the diffuse reflectance.

Forward model of optical scattering in skin

Monte Carlo simulation for light transport in multi-layered tissue.

The earlier work of Lihong Wang and Steven L. Jacques produced an excellent standard C-code for Monte Carlo simulation of light transport in multi-layered tissue (MCML)¹¹. Figure 1 is an example of how Wang and Jacques utilized such a simulation to model the propagation of a photon through a homogenous medium. Our research utilizes the MCML program to simulate the spectral reflectance of skin.

The MCML program is constituted by following operations for photons¹².

- (1) Photon launching
- (2) Generating the propagation distance

- (3) Moving the photon
- (4) Internal reflection
- (5) Photon absorption
- (6) Changing photon direction by scattering
- (7) Calculating observable quantities

The Monte Carlo simulation begins by launching a photon into the tissue. Each propagation distance between photon positions is variable and equals $\Delta s = -\ln(\xi)/(\mu_a + \mu_s)$ where ξ is a random number and μ_a and μ_s are absorption and scattering coefficients, respectively. Photon's spatial position is described by three Cartesian coordinates. The direction of photon movement is described by three directional cosines. The directional cosines are denoted by μ_x , μ_y , and μ_z corresponding to the x, y, and z axes, respectively. The new coordinates (x', y', z') are calculated for a photon at (x, y, z) to the direction (μ_x , μ_y , μ_z) as follows.

$$\begin{aligned}
 x' &= x + \mu_x \Delta s \\
 y' &= y + \mu_y \Delta s \\
 z' &= z + \mu_z \Delta s
 \end{aligned} \tag{2}$$

The probability that the photon will be internally reflected is determined by the Fresnel reflection coefficient when the photon is propagated across a boundary into a region with a different index of refraction.

A weight is assigned to each photon to improve the efficiency of a Monte Carlo programs by assuming that many photons (a packet) propagate along the same pathway. The weight of the photon is decreased

with absorption from an initial value of 1 as it moves through the tissue, and equals a^n after n steps, where a is the albedo ($a = \mu_s / (\mu_a + \mu_s)$).

A normalized phase function describes the probability density function for the azimuthal and longitudinal angles (θ, ϕ) for a photon when it is scattered. The scattering in tissue is easily characterized by the generating function for Henyey-Greenstein phase function as follows.

$$\cos \theta = \frac{1}{2g} \left\{ 1 + g^2 - \left[\frac{1 - g^2}{1 - g + 2g\xi} \right] \right\} \quad (3)$$

where ξ is a uniformly distributed random number from 0 to 2π , the value of g is called the anisotropy factor of scattering. The anisotropy characterize the type of scattering such as backward, forward, isotropic scatterings. If a photon is scattered at an angle (θ, ϕ) from the direction (μ_x, μ_y, μ_z) in which it is moving, then the new direction (μ'_x, μ'_y, μ'_z) is calculated by

$$\begin{aligned} \mu'_x &= \frac{\sin \theta}{\sqrt{1 - \mu_z^2}} (\mu_x \mu_z \cos \phi - \mu_y \sin \phi) + \mu_x \cos \theta \\ \mu'_y &= \frac{\sin \theta}{\sqrt{1 - \mu_z^2}} (\mu_y \mu_z \cos \phi - \mu_x \sin \phi) + \mu_y \cos \theta \\ \mu'_z &= -\sin \theta \cos \phi \sqrt{1 - \mu_z^2} + \mu_z \cos \theta \end{aligned} \quad (4)$$

Many photon trajectories are calculated to yield a statistical description of photon distribution in the medium. The ratio between the sum of weights that escape from the surface and sum of initial weights used in the simulation gives the diffuse reflectance.

Three-layered model of human skin and its spectral reflectance by Monte Carlo simulation.

Figure 2 shows the three-layered model of skin, which is composed of the epidermis, dermis and subcutis layers. In this model, it is supposed that the skin color is dominated by the concentration of melanin in the epidermis layer, oxy-hemoglobin and deoxy-hemoglobin in the dermis layer. The refractive index in air and tissue¹³ are set for 1.0 and 1.4 respectively. The depths of epidermis, dermis and subcutis layers are set for 0.007 cm, 0.113 cm, and 0.5 cm respectively. The depth of epidermis layer was measured by ourselves for a male forearm. At this measurement, the technique of the optical coherence tomography^{14,15} was performed by using the optical reflectometer (Hewlett-Packard Model HP 8504A). The depth of dermis layer is set for the measured mean value for a male forearm shown in the reference 16. The ultrasonic diagnostic equipment was used at the measurement. The depth of subcutis layer is set empirically to reflect the most of light onto the dermis layer. The scattering coefficient and anisotropy factor used in epidermis and dermis layers are shown in Fig. 3, which were published by Germert et al.¹⁷. The scattering coefficient and anisotropy factor in subcutis layer are set for 600 [cm⁻¹] and 0.8 empirically in all wavelengths respectively. The absorption cross sections of melanin $\mu_a^{Melanin}$, oxy-hemoglobin $\mu_a^{HbO_2}$ and deoxy-hemoglobin μ_a^{Hb} are shown in Fig. 4, which were published by Anderson et al.⁶. In this research, variable parameters in the skin model are the only concentrations of melanin $C^{Melanin}$, oxy-hemoglobin C^{HbO_2} and deoxy-hemoglobin C^{Hb} .

Using MCML program described in the previous sub-section, diffuse reflectance in this three-layered skin model can be easily simulated. It is required to run the program for each wavelength to obtain the spectral

reflectance using the spectral absorptions and scattering coefficients shown in Figs 3 and 4. For single spectral reflectance of skin, therefore, we run the MCML programs 31 times since the wavelength is sampled between 400nm and 700nm in the interval of 10nm. One million photons were used for each wavelength in the program. To evaluate the error of simulation, it was repeated 10 times changing the initial random value. Figure 5(a) shows the resultant averaged spectral reflectance for 10 times simulations, where the concentrations of melanin, oxy-hemoglobin and deoxy-hemoglobin were set to 150, 9, 1 [$\times 10^{-5}$ mol/L] respectively. The standard deviation of error in each wavelength for 10 data is shown in Fig. 5(b). It is shown that the deviation of Monte Carlo simulation is under 3.5×10^{-4} in all wavelengths.

Inverse optical scattering technique

The concentration of pigments is estimated from diffuse spectral reflectance. Figure 6 shows the schematic diagram of the estimation method. At first, the initial values of variable parameters $c^{Melanin}$, c^{HbO_2} , c^{Hb} are randomly given, and the spectral reflectance is simulated using the initial parameters based on the forward model of optical scattering in human skin. The amount of error is calculated between the simulated spectral reflectance and measured spectral reflectance. If the error is not small enough, the variable parameters are changed, and spectral reflectance is simulated again by using the changed parameters. If the error is small enough, the parameters used for the simulation are results of the estimation. To decide the change of parameters, a conventional optimization technique is used by MATLAB optimization tool box¹³. In this research, we used the reflectance on 580 nm, 610 nm and 640 nm wavelength for the estimation. Figure 7 shows

examples of the estimation. In Fig. 7(a), concentrations of melanin and total hemoglobin are set for 150×10^{-5} mol/L, 10×10^{-5} mol/L, respectively. The saturation is changed from 0.0 to 1.0. The horizontal axis indicate the target saturation value for the estimation, vertical axis indicate the estimated saturation value by the above method. In Fig. 7(b), concentrations of total hemoglobin and saturation are set for 10×10^{-5} mol/L, 0.9 respectively. The concentrations of melanin are changed from 0 to 300×10^{-5} mol/L. In Fig. 7(c), concentrations of melanin and saturation are set for 150×10^{-5} mol/L, 0.9 respectively. The concentrations of hemoglobin are changed from 0 to 30×10^{-5} mol/L. These results show that the concentrations and saturation are successfully estimated by the proposed technique.

Experiments

The imaging system consists of standard white-light source (Natural Light NL-500, USHI/MURAKAMI) and a multi-spectral camera (OLYMPUS)⁴. A wheel with 10 interference color filters is rotating between the lens and monochromatic CCD in the multi-spectral camera. The peak wavelengths of filter transmittance are arranged between 430nm~700nm in 30 nm intervals. In our experiment, only three color filters; 580, 610, 640 nm were used for the estimation of pigmentation, since the images by other filters were suffered from noise strongly in the experiment. A reference white diffuse reflectance plate was placed on the skin and was used to calibrate the imaging system. During the occlusions in the elbow by a bandage, the skin image in the forearm of subject was captured every 30 seconds. The subject is a Japanese young student in our laboratory. He was also the subject for measuring the depth of epidermis to model the skin structure. In the experiment, two types of occlusion; venous occlusion, arterial and venous occlusion, are performed as follows.

The venous occlusion was performed by applying the bandage softly in the elbow. On the other hand, the arterial and venous occlusion was performed by applying the bandage strongly. The strength of binding for both occlusions was adjusted empirically. The occlusions were continued for 3 minutes. After releasing the occlusions, the skin image was captured every 30 seconds. Figure 8 shows the results of experiment. The oxygen saturation decreases more rapidly and decreases to a lower level for venous and arterial occlusion than for venous occlusion. On the other hand, the blood volume increases more rapidly and increases to a higher level for venous occlusion than for venous and arterial occlusion. The estimated concentration of melanin was almost constant during and after the occlusions. These results agree with physiological knowledge that concentration of melanin will not change by occlusions, that concentration of hemoglobin will increase in congestion by the venous occlusion, and that oxygen saturation decrease with the change of inflow of arterial blood. Therefore, we think that the proposed technique is an effective means for estimating the 2-D distribution of the relative value of oxygen saturation, blood volume and pigmentation in the skin area. By imaging a slapped region of the human forearm, the increase of oxygen saturation and blood volume were imaged by our estimation technique as shown in Fig. 9.

Conclusion

The maps of melanin, oxy-hemoglobin and deoxy-hemoglobin in skin are estimated from multi-visible-spectral image by using the inverse optical scattering technique. Influences of noise, 3-D shape, specularity, skin types should be investigated in further investigations. However, the results show a possibility to diagnose a skin disease from multi-channel visible spectrum imaging by a collaboration with medical community,

and resultant map will give useful information in reproducing various skin colors in spectral-based color management^{3,4, 18} and computer graphics.

Acknowledgement

The authors thank Mr. Tetsuya Tsugita and Mr. Koji Minami in Kao Corporation for supporting us in measuring the depth of epidermis by using their instrument.

References

1. S. Kawata, K. Sasaki and S. Minami, Component analysis of spatial and spectral patterns in multispectral images. I. Basis, *Journal of the Optical Society of America A* 4, 2101-2106 (1987).
2. F. H. Imai, N. Tsumura, H. Haneishi and Y. Miyake, Principal component analysis of skin color and its application to colorimetric color reproduction on CRT display and hardcopy, *J. Image Science and Technology* 40, 422-430 (1996).
3. Y. Yokoyama, N. Tsumura, H. Haneishi, Y. Miyake, J. Hayashi and M. Saito, A new color management system based on human perception and its application to recording and reproduction of art printing, in *Proceedings of IS&T/SID's 5th Color Imaging Conference, Color Science, Systems and Appl.* (Society for Imaging Science & Technology, Springfield, VA, 1997) pp. 169-172 .
4. M. Yamaguchi, R. Iwama, Y. Ohya, T. Obi, N. Ohyama and Y. Komiya, Natural color reproduction in the television system for telemedicine, in *Medical Imaging 1997: Image Display*, Y. Kim, ed., *Proc. Soc. Photo-Opt. Instrum. Eng.* 3031, 482-489 (1997).

5. M. J. C. van Gemert, S.L. Jacques, H. J. C. M. Sternborg and W. M. Star, Skin optics, *IEEE Trans. on Biomedical Engineering* 36, 1146-1154 (1989).
6. R. R. Anderson and J. A. Parrish, The optics in human skin, *The Journal of Invest. Dermatology* 77, 13-19 (1981).
7. N. Tsumura, H. Haneishi and Y. Miyake, Independent component analysis of skin color image, *Journal of the Optical Society of America A* 16, pp. 2169-2176 (1999).
8. N. Tsumura, H. Haneishi and Y. Miyake, Independent component analysis of skin absorbance image, *Optical Review* 7(6), 479-482(2000).
9. P. J. Dwyer, R. R. Anderson and C. A. DiMarzio, Mapping blood oxygen saturation using a multi-spectral imaging system, in *Biomedical Sensing, Imaging, and Tracking Technologies II*, T. Vo-Dinh, R. A. Lieberman, and G. G. Vurek, eds., *Proc. Soc. Photo-Opt. Instrum. Eng.* 2976, 270-280 (1997).
10. I. Yoshida, Y. Shimada, K. Tanaka, Spectrophotometric monitoring of arterial oxygen saturation in the fingertip, *Med. & Biol. Eng., Comput.*, Vol. 18, 27-32(1980).
11. Wang, L., Jacques, S. L., Monte Carlo Modeling of Light Transport in Multi-layered Tissues in Standard C, University of Texas M. D. Anderson Cancer Center(1992).

Web page <http://omlc.ogi.edu/software/mc/index.html>

12. Prahl, S.A., M. Keijzer, S.L. Jacques, A.J. Welch: A Monte Carlo model of light propagation in tissue. In *Dosimetry of Laser Radiation in Medicine and Biology*, G. Mueller, D. Sliney, Eds., SPIE Series Vol. IS 5:102-111 (1989).
13. Frank P. Bolin, Luther E. Preuss, Roy C. Taylor, and Robert J. Ference, Refractive index of some mammalian tissues using a fiber optic cladding method, *Appl. Opt.* Vol. 28, No.12, pp.2297-2303 (1989).

14. D. Huang, E. A. Swanson, C. P. Lin, J. S. Schuman, W. G. Stinson, W. Chang, M. R. Hee, T. Flotte, K. Gregory, C. A. Puliafito, J. G. Fujimoto, Optical Coherence Tomography, *Science*, 254, 1178 (1991).
15. J. M. Schmitt, A. Knüttel, and R. F. Bonner, Measurement of optical properties of biological tissues by low-coherence reflectometry, *Appl. Opt.* Vol. 32, No. 30, pp. 6032-6042 1993.
16. Michiro Shimozuma, Hidemi Nakagawa, Fujio Otsuka and Yasumasa Ishibashi, Application of Ultrasonic Diagnostic Equipment in Dermatology - Measurement of Skin Thickness -, *The Japanese journal of dermatology*, Vol. 98, No. 1, 65-68 (1988). [in Japanese]
17. Gemert, M.J.C.V., Jacques, S. L., Sternborg, H. J. C. M. and Star, W.M., Skin optics, *IEEE Trans on Biomed. Eng.* 36, 1146 (1989).
18. M. R. Rosen, M. D. Fairchild, G. M. Johnson, D. R. Wyble, Color Management within a Spectral Image Visualization Tool, in *Proceedings of IS&T/SID's 8th Color Imaging Conference, Color Science, Systems and Appl.* (Society for Imaging Science & Technology, Springfield, VA, 2000) pp. 75-80 .

Figure captions

Figure 1 The movement of one photon through a homogenous medium, as calculated by Monte Carlo simulation.

Figure 2 Three layered model of skin.

Figure 3 Scattering coefficient and anisotropy factor in epidermis and dermis layer.

Figure 4 Absorption cross section of melanin, oxy-hemoglobin (HbO₂), and deoxy-hemoglobin (Hb).

Figure 5(a) Result of Monte Carlo simulation and (b) Standard deviation of error in the simulation.

Figure 6 Estimation of concentrations of pigments from diffuse spectral reflectance.

Figure 7 Results of estimation of concentrations of pigments from absolute diffuse spectral reflectance.

Figure 8 Results of estimation of concentrations of pigments from imaging the human forearm under (a) venous and arterial occlusion, (b) venous occlusion.

Figure 9 Results of estimation of concentrations of pigments from imaging the slapped region of the human skin.